



Certificate of Compliance

Certificate: 2700973

Master Contract: 171882

Project: 70043941

Date Issued: October 29, 2015

Issued to: Honeywell Analytics Ltd.
Hatch Pond House
4 Stinsford Rd
POOLE
Dorset
BH17 0RZ
UNITED KINGDOM

Attention: Mr. Colin Hydes

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only



Issued by:

A handwritten signature in black ink, appearing to read 'Steven Collins'.

Steven Collins

PRODUCTS

CLASS 2252 05 - PROCESS CONTROL EQUIPMENT

CLASS 2252 85 - PROCESS CONTROL EQUIPMENT (Certified to U.S. Standards)

Touchpoint Pro (TPPR) Controller, OEM Unit 5U assembly, permanently connected

Rated:

Input: 24 V_{DC} Nominal (18-32V_{DC}), Relay Contacts: 30V_{DC}/ 250V_{ac}, 5A, Power Consumption: 13W max

Notes:

- The above model is Equipment Class I grounded, Pollution Degree 2.
- Mode of operation: Continuous.
- Environmental Conditions: Extended: -20°C to 65°C, 2000 m max (Altitude), 10 to 90%RH (non-condensing).
- 19" Rack 5U Assembly is alternate name of OEM Unit 5U when mounted in Honeywell 5 U 19 open rack unit.



Certificate: 2700973
Project: 70043941

Master Contract: 171882
Date Issued: October 29, 2015

CONDITIONS OF ACCEPTABILITY

1. Equipment is to be installed by manufacturer trained personnel or competent person trained in accordance with manufacturer installation instructions.

APPLICABLE REQUIREMENTS

- CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements
- UL Std. No. 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements



Supplement to Certificate of Compliance

Certificate: 2700973

Master Contract: 171882

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
70043941	October 29, 2015	Change of material used as an overlay on the front panel of the equipment.
70032844	May 13, 2015	Report update to include Honeywell 5 U 19" Powder coated Mild Steel open frame rack unit as an alternate controller mounting configuration.
2700973	September 26, 2014	Original CSA C-US Certification according to CAN/CSA C22.2 No. 61010-1-04 / UL 61010-1, 2 nd Edition.