



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

NIOSH Reference: TN-12162
Mfr. Reference: SUR01J13A

Centers for Disease Control
and Prevention (CDC)
National Institute for Occupational
Safety and Health - ALOSH
1095 Willowdale Road
Morgantown, WV 26505-2888

Phone: (304) 285-5907
Fax: (304) 285-6030
January 8, 2002

Mr. Nick Luzie
Project Engineer
Survivair, Inc.
3001 S. Susan Street
Santa Ana, California 92704

Dear Mr. Luzie:

The National Institute for Occupational Safety and Health (NIOSH) has reviewed your request dated September 13, 2001, and received September 26, 2001. This request was for an extension of approval as listed on assembly matrix N9684AML.XLS for the following changes:

1. To add three new hood-style facepieces part numbers (968005, 968006, and 968007) for use with the Survivair Hip-Pac SCBA/SAR series respirators TC-13F-300, TC-13F-301, TC-13F-367, and TC-13F-482, and the Survivair Hip-Pac SAR respirator TC-19C-302. The hood-style facepieces and pneumatics are currently approved for use under TC-13F-447.
2. To add a new anti-fog wipe part number 981805 for use with the Survivair Hip-Pac SCBA/SAR series respirators TC-13F-300, TC-13F-301, TC-13F-367, and TC-13F-482 and the Survivair Hip-Pac SAR respirator TC-19C-302, and the Survivair Hip-Pac SAR/P100 respirators TC-84A-229.
3. To add several facepieces and accessories to approvals TC-13F-300, TC-13F-301, TC-13F-367, and TC-19C-302, and TC-84A-229. These facepieces and accessories have all been previously approved on the 4500 PSI Survivair Hip-Pac SCBA/SAR under TC-13F-482.
4. To remove the words "buddy breather" from accessory item 9AA on the assembly matrix and the drawing for this item.

This extension of approvals is granted. Approval is granted for English language only on all documentation. It is the manufacturer's responsibility to correctly translate materials desired in languages other than English.

Page 2 - Mr. Nick Luzie - TN-12162

The diskettes enclosed with this letter contain the final respirator approval labels for this submission. The cautions and limitations which apply to these approvals are on the approval labels. Only those assemblies where approval numbers are assigned apply to these approval actions. Production approval labels cannot include information on unapproved configurations.

The approved assembly consists of the parts as listed on the approval labels plus the assembly matrix. Parts are to be marked with the numbers indicated on the approval label in a legible and permanent manner (marking cannot be removed without evidence of its previous presence).

No additional changes may be made to any respirators and accompanying documentation without prior written approval of NIOSH. Requests for changes must be submitted to NIOSH and a modification of this approval must be granted before any changes are made.

Sincerely yours,



RW Richard W. Metzler
Acting Director
National Personal Protective Technology Laboratory

Enclosure