

TECHNICAL DATA

North BP1000 Backpack Adapter

NIOSH Approval Information

The National Institute of Occupational Safety and Health (NIOSH) has made several changes to the Tested and Certified (TC) designations for all approvals for the North Backpack Adapter, BP1002 when used with North half masks and full facepieces. The reason for the change is that NIOSH now requires separate TC numbers for facepiece combinations with and without the backpack adapter. Product already in use is certified with the new TC numbers, when used in any of the configurations below.

The approval numbers for the RU8500 with the Backpack adapter have not changed at this point, but will also have new approval numbers in the future.

Note that the RU6500 is not approved with the BP1002 Backpack adapter.

For more information and copies of the NIOSH Approval Labels please contact Technical Support

Phone: 800.873.5242, select 2

email: hsptechsupport@honeywell.com



BP1000 Shown with 7700 Half Mask



BP1000 Backpack Adapter

NORTH BP1000 BACKPACK ADAPTER CERTIFICATIONS

FILTER & CARTRIDGE USED		FACEPIECES			
		5500	7700	5400	7600
7580P100	P100	TC-84A-7608	TC-84A-7607	TC-84A-7626	TC-84A-7621
7581P100L	OV/P100	TC-84A-7613	TC-84A-7615	TC-84A-7627	TC-84A-7622
75SCP100L	Multi/P100	TC-84A-7614	TC-84A-7616	TC-84A-7628	TC-84A-7623
N75001L, 7506N95 and N750036*	OV/N95	TC-84A-7611	TC-84A-7612	TC-84A-7629	TC-84A-7624
75SCL, 7506N95 and N750036*	Multi/N95	TC-84A-7609	TC-84A-7610	TC-84A-7630	TC-84A-7625

* N750036 includes 1 pair each N750015 filter holder, N750027 filter and 750038 filter supports

For more information

www.honeywellsafety.com

Honeywell Safety Products

USA

900 Douglas Pike
Smithfield, RI 02917
Phone: 800.873.5242
www.honeywell.com

CANADA

10550 Parkway Boulevard
Anjou, Quebec H2W 2R2
Phone: 800.873.5242