



Certificate of Compliance

Certificate: 2138307 (80259)

Master Contract: 157918

Project: 2690149

Date Issued: January 16, 2014

Issued to: BW Technologies by Honeywell

2840 2nd Ave SE
Calgary, AB T2A 7X9
Canada
Attention: Kevin Giddings

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Zahra Amini

Issued by: Zahra Amini

PRODUCTS

CLASS 4828 01 - SIGNAL APPLIANCES - - Combustible Gas Detection Instruments-For Hazardous Locations

CLASS 4828 81 - SIGNAL APPLIANCES-Combustible Gas Detection Instruments - For Hazardous Location-Certified to U.S. Standards

CLASS 4828 82 - SIGNAL APPLIANCES-Toxic Gas Detection Instruments - For Hazardous Locations. Certified to U.S. Standards

CLASS 4828 02 - SIGNAL APPLIANCES - Toxic Gas Detection Instruments - For Hazardous Locations

Class I, Division 1, Groups A,B,C,D

Gas Alert Quattro Portable Gas Detector, QT-XWHM-A-Y-...; Temperature Code T4 or T135.3°C @ -20°C to +50°C ambient; Intrinsically Safe when powered by Lithium Polymer battery pack part number QT-BAT-Rxxx or "AA" Alkaline Cells (Duracell MN1500, Energizer E91VP)

APPLICABLE REQUIREMENTS



Certificate: 2138307 (80259)

Master Contract: 157918

Project: 2690149

Date Issued: January 16, 2014

C22.2 No 0 - M1991 - General Requirements - Canadian Electrical Code Part II.

C22.2 No 152 - M1984 - Combustible Gas Detection Instruments.

C22.2 No 157 - M1992 - Intrinsically Safe and Non-Incendive Equipment for Use in Hazardous Locations.

UL 913, Seventh Edition - Intrinsically Safe Apparatus and Associated Apparatus for use in Class I, II, III, Division 1, Hazardous (Classified) Locations.

ANSI / ISA 12.13.01 – 2000 – Performance Requirements for Combustible Gas Detectors