



# Certificate of Compliance

**Certificate:** 2700973

**Master Contract:** 171882

**Project:** 70043941

**Date Issued:** October 29, 2015

**Issued to:** Honeywell Analytics Ltd.  
Hatch Pond House  
4 Stinsford Rd  
POOLE  
Dorset  
BH17 0RZ  
UNITED KINGDOM

**Attention:** Mr. Colin Hydes

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only*



**Issued by:**

A handwritten signature in black ink, appearing to read 'Steven Collins'.

Steven Collins

## **PRODUCTS**

**CLASS 2252 05 - PROCESS CONTROL EQUIPMENT**

**CLASS 2252 85 - PROCESS CONTROL EQUIPMENT (Certified to U.S. Standards)**

Touchpoint Pro (TPPR) Controller, OEM Unit 5U assembly, permanently connected

### **Rated:**

Input: 24 V<sub>DC</sub> Nominal (18-32V<sub>DC</sub>), Relay Contacts: 30V<sub>DC</sub>/ 250V<sub>ac</sub>, 5A, Power Consumption: 13W max

### **Notes:**

- a. The above model is Equipment Class I grounded, Pollution Degree 2.
- b. Mode of operation: Continuous.
- c. Environmental Conditions: Extended: -20°C to 65°C, 2000 m max (Altitude), 10 to 90%RH (non-condensing).
- d. 19" Rack 5U Assembly is alternate name of OEM Unit 5U when mounted in Honeywell 5 U 19 open rack unit.



**Certificate:** 2700973  
**Project:** 70043941

**Master Contract:** 171882  
**Date Issued:** October 29, 2015

---

**CONDITIONS OF ACCEPTABILITY**

1. Equipment is to be installed by manufacturer trained personnel or competent person trained in accordance with manufacturer installation instructions.

**APPLICABLE REQUIREMENTS**

- CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements
- UL Std. No. 61010-1 (2<sup>nd</sup> Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements



## *Supplement to Certificate of Compliance*

**Certificate:** 2700973

**Master Contract:** 171882

*The products listed, including the latest revision described below,  
are eligible to be marked in accordance with the referenced Certificate.*

### **Product Certification History**

---

<b>Project</b>	<b>Date</b>	<b>Description</b>
70043941	October 29, 2015	Change of material used as an overlay on the front panel of the equipment.
70032844	May 13, 2015	Report update to include Honeywell 5 U 19" Powder coated Mild Steel open frame rack unit as an alternate controller mounting configuration.
2700973	September 26, 2014	Original CSA C-US Certification according to CAN/CSA C22.2 No. 61010-1-04 / UL 61010-1, 2 <sup>nd</sup> Edition.