

Product Data Sheet

Product Datasheet 3FF Carbon Monoxide Sensor

Document Purpose

The purpose of this document is to present the performance specification of the 3FF Carbon Monoxide gas sensor.

This document should be used in conjunction with the Operating Principles (OP08) and the Product Safety Datasheet (PSDS 16).

The data provided in this document are valid at 20°C, 50% RH and 1013 mBar for 3 months from the date of sensor manufacture.

Output signal can drift below the lower limit over time. For guidance on the safe use of the sensor, please refer to the Operating Principles OP08.

Product Data Sheet

- Key Features & Benefits:**
- Robust, 3-Series packaging
 - Range of accessories available

Technical Specifications

MEASUREMENT

Operating Principle	3-electrode electrochemical
Measurement Range	0-4000 ppm CO
Maximum Overload	20000 ppm CO
Filter	To remove acid gases
Filter Life	560,000 ppm hours <small>See note 1</small>
Sensitivity*	0.030 ± 0.006 µA/ppm
Resolution	1 ppm CO
Response Time (T₉₀)*	<30 seconds
Baseline Offset (clean air)	-3 to +10 ppm equivalent
Zero Shift (+20°C to +40°C)	<20 ppm equivalent
Repeatability	1% of signal
Linearity	Linear

ELECTRICAL

Recommended Load Resistor	10 Ω
Bias Voltage	Not Required

MECHANICAL

Weight	22 g
Orientation	Any

ENVIRONMENTAL


Operating Temperature Range	-20°C to +50°C
Recommended Storage Temp	0°C to 20°C
Operating Pressure Range	Atmospheric ± 10%
Pressure Coefficient	0.007 ± 0.003 % signal/mBar
Operating Humidity Range	15 - 90% RH non-condensing

LIFETIME

Long Term Sensitivity Drift*	<2% signal loss/month
Expected Operating Life	Three years in air
Storage Life	6 months in CTL container

Note 1 : NO removal based on continuous exposure to 1000 ppm and 5% breakthrough

Product Dimensions



All dimensions in mm

All tolerances ±0.15 mm unless otherwise stated
3FF shown with side tags and tin pins

AVAILABLE OPTIONS

Sensor	Description	Part Number
3FF	With side tag and PCB pin connections	ABG06-J08
3FF(S)	With side tag connection	ABG06-008
3FF(G)	With gold-plated PCB pin connection	ABG06-308

IMPORTANT NOTE:

Soldering to the pin connections will seriously damage the sensor and invalidate the warranty.

* Specifications are valid at 20°C, 50% RH and 1013 mBar, using City Technology recommended circuitry. Performance characteristics outline the performance of sensors supplied within the first 3 months. Output signal can drift below the lower limit over time.

Product Data Sheet

Poisoning

CiTiceLs are designed for operation in a wide range of environments and harsh conditions. However, it is important that exposure to high concentrations of solvent vapours is avoided, both during storage, fitting into instruments and operation.

When using sensors with printed circuit boards (PCBs), degreasing agents should be used before the sensor is fitted. Do not glue directly on or near the CiTiceL as the solvent may cause crazing of the plastic.

Cross Sensitivity Table

Whilst CiTiceLs are designed to be highly specific to the gas they are intended to measure, they will still respond to some degree to various other gases. The table below is not exclusive and other gases not included in the table may still cause a sensor to react.

IMPORTANT NOTE : The cross sensitivity data shown below does not form part of the product specification and is supplied for guidance only. Values quoted are based on tests conducted on a small number of sensors and any batch may show significant variation. For the most accurate measurements, an instrument should be calibrated using the gas under investigation.

Gas	Concentration Used	3FF (ppm CO)
Hydrogen Sulfide, H ₂ S	15	0
Sulfur Dioxide, SO ₂	5	0
Nitric Oxide, NO	50	-1 < x\$ < 0
Nitrogen Dioxide, NO ₂	50	-1 < x\$ < 0
Hydrogen, H ₂	100	<60 <small>See note 2</small>
Hydrogen Chloride, HCl	5	0
Ethylene, C ₂ H ₄	100	0 < x\$ < 20

Note 2 : For applications where a hydrogen compensated output is required, the A3ED should be used

SAFETY NOTE

This sensor is designed to be used in safety critical applications. To ensure that the sensor and/or instrument in which it is used, are operating properly, it is a requirement that the function of the device is confirmed by exposure to target gas (bump check) before each use of the sensor and/or instrument. Failure to carry out such tests may jeopardize the safety of people and property.

Every effort has been made to ensure the accuracy of this document at the time of printing. In accordance with the company's policy of continued product improvement City Technology Limited reserves the right to make product changes without notice. The products are always subject to a programme of improvement and testing which may result in some changes in the characteristics quoted. As the products may be used by the client in circumstances beyond the knowledge and control of City Technology Limited, we cannot give any warranty as to the relevance of these particulars to an application. It is the clients' responsibility to carry out the necessary tests to determine the usefulness of the products and to ensure their safety of operation in a particular application.