



(H) SPECIFICATION No. -S1600

CAPACITY _____ (A) 2000, 5000 & 10,000 IN.-LBS.
 O'LOAD CAPACITY _____ 300% @ 2K & 5K IN.-LBS. & 200% @ 10K IN.-LBS.
 SPEED _____ 12,000 R.P.M. (I)
 OUTPUT @ RATED CAP _____ 2.0 MV/VOLT NOM.

NOTE:
 CONSULT INSTRUCTION MANUAL BEFORE
 ATTEMPTING DISASSEMBLY OF UNIT.

I	RPM WAS 15000	2/2/80
H	ECN 2086	3/9/80
G	ADD SPEC. NO.	8/28/80
F	PATENT TAG & WARNING LABEL WERE NOT SHOWN	3-18-81
E	ECN 2702 CORRECTED	9-19-78
D	ECN 2702	5/25/78
C	ADDED DIM. M.L.	8/26/77
B	ECN 2116 R.T.P. R.Z.	11-7-72
A	2K CAP. ADDED	7-27-72

DET. NO.	PART NO.	DESCRIPTION	REQ'D
		INSTALLATION DWG ~	
		ROTARY TRANSFORMER TORQUE SENSORS	
SCALE:	DRAWN BY:	REF. DWG:	MAT'L:
FULL	KA	C-1957B	
DATE:	CHK'D BY:	MODEL:	JOB NO.
4-18-72	KA	1805	

UNLESS OTHERWISE SPECIFIED:
 TOLERANCE ON FRACTIONS ± 1/64" TOLERANCE ON DRILLED HOLES ± .007
 TWO PLACE DECIMALS ± .010 REMOVE ALL BURRS & SHARP EDGES
 THREE PLACE DECIMALS ± .005 DIMENSION LIMITS HELD AFTER PLATING
 ANGLES ± 0'30" SURFACE FINISH 125.7