



(H) SPECIFICATION No. -S1600

CAPACITY _____ (A) 2000, 5000 & 10,000 IN.-LBS.
 O'LOAD CAPACITY _____ 300% @ 2K & 5K IN.-LBS. & 200% @ 10K IN.-LBS.
 SPEED _____ 12,000 R.P.M. (I)
 OUTPUT @ RATED CAP _____ 2.0 MV/VOLT NOM.

NOTE:
 CONSULT INSTRUCTION MANUAL BEFORE
 ATTEMPTING DISASSEMBLY OF UNIT.

REV	DESCRIPTION	DATE	BY	CHK'D BY	DATE	SCALE	REF. DWG.	MAT'L	JOB NO.
I	RPM WAS 15000 ECN 2086	2/2/80	ame						
H	ADD SPEC. NO.	3/9/80							
G	PATENT TAG & WARNING LABEL WERE NOT SHOWN REV 3007	8/28/80							
F	ECN 2970	3/18/81							
E	ECN 2702 CORRECTED	9/19/78							
D	ECN 2702	5/25/78							
C	ADDED DIM m.l.	8/26/77							
B	ECN 2116 R.T.P. R.Z.	11-7-72							
A	2K CAP. ADDED	7-27-72							

DET. NO.	PART NO.	DESCRIPTION	REQ'D
		INSTALLATION DWG ~	
		ROTARY TRANSFORMER TORQUE SENSORS	

SCALE	DRAWN BY	REF. DWG.	MAT'L
FULL	KA	C-1857B	

DATE	CHK'D BY	MODEL	JOB NO.
4-18-72	KA	1805	

UNLESS OTHERWISE SPECIFIED:	
TOLERANCE ON FRACTIONS	± 1/64 TOLERANCE ON DRILLED HOLES ± .007
TWO PLACE DECIMALS	± .010 REMOVE ALL BURRS & SHARP EDGES
THREE PLACE DECIMALS	± .005 DIMENSION LIMITS HELD AFTER PLATING
ANGLES ± 0'30"	SURFACE FINISH 125.7

SYM.	REVISION	DATE
J		

LEBOW ASSOCIATES, INC.
 OAK PARK, MICHIGAN

DRAWING NO. C-1605-D